

ERRATA

Raphex 2012 Therapy Answer Book

Page 3, answer T22 The correct answer is “C.” Replace the answer with the following:

- C Most port films are taken with 6 MV photon beams, for which over 95% of interactions are Compton. There is very little pair production or photoelectric, but slightly more pair production than photoelectric production.

Page 4, answer T39 The correct answer, as listed in the exam booklet, is:

- D Total number of monitor units is an important part of plan checking but is not a measure of plan quality.

(It has been noted that option “C. Beam arrangement” could also be correct since if a plan had too many monitor units, then there would be an increased risk of induced cancer. *Therefore, question T39 was omitted in grading the exam.*)